


Detlef Janke 
Galvanik-Galvanoformung-Spritzlackiererei
Hohentwielsteig 3
14163 Berlin

Fischerscope® XRAY XAN 222
Product: 7910021 / Au/Ni/Cu Dir.: Fischer Block: 24
Kunde: Trudika-Shop Artikel : AHG200707002



n=	1	Au 1 =	5.18 μm	Ni 2 =	4.41 μm
n=	2	Au 1 =	5.18 μm	Ni 2 =	4.54 μm
n=	3	Au 1 =	5.17 μm	Ni 2 =	4.44 μm
n=	4	Au 1 =	5.26 μm	Ni 2 =	6.25 μm
n=	5	Au 1 =	5.25 μm	Ni 2 =	6.27 μm
n=	6	Au 1 =	5.27 μm	Ni 2 =	6.14 μm

Mean	5.22 μm	5.34 μm
Standard deviation	0.045 μm	0.964 μm
C.O.V. (%)	0.86	18.05
Range	0.10 μm	1.86 μm
Number of readings	6	6
Min. reading	5.17 μm	4.41 μm
Max. reading	5.27 μm	6.27 μm
Measuring time	30 sec	
Operator: DJ	Trudika-Shop	AHG200707002
Date: 01.08.2020 Time: 13:15:13		